

1. PRODUCT IDENTIFIER

Code DE7430
Name Insulin Ab ELISA

Insulin Ab ELISA is a medical device for the quantitative detection of IgG class autoantibodies against Insulin in human serum or plasma. Intended for professional in vitro diagnostic use only.

Manufacturer / Supplier

Demeditec Diagnostics GmbH
Lise-Meitner-Str. 2
24145 Kiel, Germany
Phone: +49 431 / 71922 0
Fax: +49 431 / 71922 55
info@demeditec.de
www.demeditec.de

2. HAZARDS IDENTIFICATION

Product is **not classified as hazardous** according to the European Regulation 1999/45/EC or 1272/2008/EC. Human health hazards: No specific hazard.

3. COMPOSITION / INFORMATION ON INGREDIENTS

Coated microplate: Purified antigen coated onto polystyrene microwells.
Calibrators/ controls: Human antibodies < 1% in phosphate buffered saline with Tween 20 as detergent, sodium azide 0.09% as preservative and BSA for stabilization.
Enzyme conjugate: Peroxidase conjugated anti-human antibody < 0.0001% in phosphate buffered saline with ProClin 300 as preservative and bovine serum albumin (BSA) for stabilization.
Sample buffer: Phosphate buffered saline with sodium azide 0.09% as preservative and BSA for stabilization.
Wash buffer: Tris buffered saline with Tween 20 as detergent and sodium azide 0.09% as preservative
Substrate solution: Aqueous solution of TMB (3,3',5,5'-Tetramethylbenzidin) 0.032% with organic solvent 2-pyrrolidone < 10%, sodium perborate, citrate, EDTA, and Kathon CG as preservative
Stop solution: Aqueous solution of phosphoric acid 4.5%.
Active substances in all mixtures do not meet the criteria for classification according to 1272/2008/EC.

4. FIRST AID MEASURES

Skin Contact: In case of skin contact, immediately wash thoroughly with water and soap. Remove contaminated clothing and shoes and wash before reuse. If stop solution comes into contact with skin, wash thoroughly with water.
Eye Contact: After contact with the eyes carefully rinse the opened eye with running water for at least 10 minutes. Get medical attention if necessary. Remove contact lenses if this can be done easily.
Respiratory tract: Take person to the fresh air.
Swallowing: Rinse the mouth and spit the fluids out. Drink 1 - 2 glasses of water immediately. During spontaneous vomiting hold the head of the casualty low with the body in a prone position in order to avoid the penetration of vomit into the air tube.

5. FIREFIGHTING MEASURES

Extinguishing Media: Use dry chemical powder, water spray, foam or carbon dioxide.

6. PERSONAL PRECAUTIONS, PROTECTIVE EQUIPMENT AND EMERGENCY PROCEDURES

Observe laboratory safety regulations.
Avoid contact with skin and eyes. Do not swallow. Do not pipette by mouth.
Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled.
When spilled, absorb with an inert material and put the spilled material in an appropriate waste disposal.

7. HANDLING AND STORAGE

Handling: Special measures are not required.
Storage: Store at 2 to 8 °C. Protect from light.

8. EXPOSURE CONTROLS / PERSONAL PROTECTION

Respiratory protection: Not required
Hand protection: Wear protective gloves of nitrile rubber or natural latex.
Eye protection: Wear protective glasses.

9. PHYSICAL AND CHEMICAL PROPERTIES

Coated microplate: Polystyrol microwells in foil pouch.
Calibrators: Yellow fluid in glass bottle.
Enzyme conjugate: Red fluid in polyethylene bottle.
Sample buffer: Yellow fluid in polyethylene bottle.
Wash buffer: Colorless fluid in polyethylene bottle.
Substrate solution: Colorless fluid in polyethylene bottle.
Stop solution: Colorless fluid in polyethylene bottle.

10. STABILITY AND REACTIVITY

Stability of components is given on the labels. Used according to intended use no dangerous reactions known.

Conditions to avoid: Substrate solution is light-sensitive. Store substrate solution in the dark.

11. TOXICOLOGICAL INFORMATION

Used according to intended use no toxicological reactions known.

12. ECOLOGICAL INFORMATION

Used according to intended use no ecological reactions known.

13. DISPOSAL CONSIDERATIONS

Waste should be disposed of in accordance with federal, state and local environmental control regulations. When disposing of conjugate solution, sample buffer or wash buffer flush drains with copious amounts of water. Disposal of packaging according to the instructions of the public authorities.

14. TRANSPORT INFORMATION

This product is not subject to official transport regulations.

15. REGULATORY INFORMATION

1907/2006/EC Registration, evaluation and authorization of chemicals regulation (REACH)
1272/2008/EC Classification, labelling and packaging regulation (CLP, globally harmonized system GHS)
replaces 67/548/EWG and 1999/45/EG, amending 1907/2006/EG
453/2010/EC Compilation of safety data sheets regulation (SDS), amending 1907/2006/EC
This product is not classified according to the EU regulations 1272/2008. No labeling requirement.

16. OTHER INFORMATION

Safety data for product including all components. This product is intended for professional laboratory use only.

Notice to Reader

To the best of our knowledge, the information contained herein is accurate. However, neither the above named supplier assumes any liability whatsoever for the accuracy or completeness of the information contained herein. All materials may present unknown hazards and should be used with caution. Although certain hazards are described herein, we cannot guarantee that these are the only hazards that exist.