


Product information

Information about other products is available at: www.demeditec.com



Ovalbumin ELISA

REF

DEOVAE01



96



Demeditec Diagnostics GmbH
Lise-Meitner-Strasse 2
24145 Kiel – Germany
www.demeditec.com

CONTENTS/ INHALTSVERZEICHNIS

1. GENERAL INFORMATION	3
2. PRINCIPLE OF THE TEST	3
3. PRECAUTIONS.....	3
4. HEALTH AND SAFETY INSTRUCTIONS.....	3
5. REAGENTS	4
6. ADDITIONAL INSTRUMENTATION AND REAGENTS (NOT PROVIDED).....	4
7. SAMPLE PREPARATION	4
8. PROCEDURE	5
9. CALCULATION OF RESULTS.....	5
10. TYPICAL STANDARD VALUES.....	5
11. PERFORMANCE	5
12. REFERENCES	7
 1. ALLGEMEINES	8
2. TESTPRINZIP	8
3. VORSICHTSMAßNAHMEN	8
4. GESUNDHEITS- UND SICHERHEITSVORSCHRIFTEN.....	8
5. REAGENZIEN	9
6. ZUSÄTZLICH BENÖTIGTE GERÄTE UND REAGENZIEN	9
7. PROBENVORBEREITUNG.....	10
8. TESTDURCHFÜHRUNG	10
9. BERECHNUNG DER ERGEBNISSE	10
10. TYPISCHE STANDARDKURVE	11
11. TECHNISCHE DATEN	11
12. LITERATUR.....	12
 SYMBOLS USED WITH DEMEDITEC ASSAYS	12

Sensitivity (Ovalbumin)	4 ppb
Recovery	95 - 115%
Incubation Time	60 min

1. GENERAL INFORMATION

Hen's egg (*Gallus gallus*) is very rich of proteins and represents an important food source for humans. While proteins of egg yolk only have minor allergenicity, many proteins of egg white are known to be allergenic. In addition to lysozyme, ovomucoid ovotransferrin and livetin, ovalbumin represents the main fraction of the egg white allergens. Amongst others egg powder as well as pure ovalbumin is often used as fining reagent for wine. For allergic persons the consumption of ovalbumin represents a critical problem. Already very low amounts of the allergen can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, ovalbumin allergic persons must strictly avoid the consumption of ovalbumin containing food. Non-declared addition of ovalbumin in food is hazardous for allergic people. Cross-contamination, mostly in consequence of the production process, is often noticed. Since July 1, 2012 the European Union requests allergen labeling for wine if milk or egg proteins are used during the production and are still present at a detection level of 0.25 mg/L or greater. Thus for the detection of ovalbumin residues, sensitive assay systems are required.

The **Demeditec Ovalbumin ELISA** represents a highly sensitive detection system for ovalbumin and is particularly capable of the quantification of ovalbumin residues in wine.

2. PRINCIPLE OF THE TEST

The **Demeditec Ovalbumin** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against ovalbumin is bound on the surface of a microtiter plate. Ovalbumin-containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against ovalbumin is given into the wells, and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of ovalbumin is directly proportional to the colour intensity of the test sample.

3. PRECAUTIONS

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipettes, ELISA reader etc.).

4. HEALTH AND SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

5. REAGENTS

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. **SORB MT** Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with anti-ovalbumin antibodies.
2. **CAL 1 - 5** Ovalbumin standards (0; 25; 100; 250; 500 ppb of ovalbumin): 5 vials with 2.0 mL each, dyed red, ready-to-use.
3. **ENZ CONJ** Conjugate (anti-ovalbumin-peroxidase): 15 mL, dyed red, ready-to-use.
4. **SUB TMB** Substrate Solution (TMB): 15 mL, ready-to-use.
5. **STOP SOLN** Stop Solution (0.5 M H₂SO₄): 15 mL, ready-to-use.
6. **SAM DIL 10x** Extraction and sample dilution buffer (Tris): 2 x 120 mL as 10x concentrate, dyed red. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. **WASH SOLN 10x** Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Instruction Manual.

6. ADDITIONAL INSTRUMENTATION AND REAGENTS (NOT PROVIDED)

Instrumentation

- 100 - 1000 µL micropipettes
- Volumetric flask
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)
- Plastic bag to store unused microtiter strips

Reagents

- double distilled water

7. SAMPLE PREPARATION

Due to high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be **cleaned thoroughly** before and after each sample. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for wine samples:

1. 1 ml of the wine sample is diluted in 19 mL of **pre-diluted** extraction buffer.
2. 100 µL of the wine solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

The following sample preparation should be applied for all other kinds of samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 1 g of the homogenized mixture is suspended in 20 mL of **pre-diluted** extraction buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

8. PROCEDURE

The washing solution is supplied as 10x concentrate and has to be **diluted 1+9** with double distilled water before use. In any case the **ready-to-use** standards provided should be determined twofold. When samples in great quantities are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation. In consideration of GLP and quality control requirements a duplicate measurement of samples is re-recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 µL **ready-to-use** standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.
5. Pipet 100 µL of conjugate (anti-ovalbumin-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 µL of stop solution (0.5 M H₂SO₄) into each well. The blue colour will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

9. CALCULATION OF RESULTS

The ready-to-use standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppb on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of ovalbumin in ppb from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

10. TYPICAL STANDARD VALUES

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 500 ppb standard. These values are only an example and should not be used instead of the standard curve which has to be measured in each new test.

Ovalbumin (ppb)	% binding of 500 ppb
500	100
250	55
100	26
25	9
0	1

11. PERFORMANCE

Sensitivity

The limit of detection (LOD) of the **Demeditec Ovalbumin test** is 4 ppb.

The limit of quantification (LOQ) of the **Demeditec Ovalbumin test** is 25 ppb.

As matrices can have variable influence on the LOD in specific cases, and the range of matrices that was tested is of course limited, the end user if needed may evaluate its own LOD values depending on the matrices to be analyzed.

Alternatively, any results below LOQ should be just reported quantitatively as "< LOQ".

Cross-reactivity

For the following foods no cross-reactivity could be detected:

Adzuki bean	Cayenne	Dill	Horseradish	Peach	Saccharose
Almond	Celery	Duck	Isinglass	Peanut	Sesame
Apricot	Cherry	Fennel	Kidney bean	Pecan	Sheep's milk
Barley	Chestnut	Fenugreek	Kiwi	Pepper, black	Shrimps
Bean, white	Chia	Fish gelatin	Lamb	Pine seed	Soy flour
Beef, cooked	Chicken	Flaxseed	Leek	Pistachio	Soy lecithin
Beef, raw	Chickpea	Garden cress	Lentil	Poppy	Split pea
Bovine gelatin	Chili	Garlic, fresh	Lupin	Pork	Sunflower seed
Brazil nut	Cinnamon	Garlic, granulated	Macadamia	Potato	Thyme
Buckwheat	Clove	Ginger, ground	Mustard, yellow	Prawn cooked	Tofu
Cabbage, white	Cocoa	Ginger, fresh	Non Fat dry milk	Prawn, raw	Tomato
Caraway seeds	Coconut	Gliadin	Nutmeg	Pumpkin seed	Turkey
Cardamom	Cod	Goat's milk	Oats	Radish	Turmeric
Carob gum	Corn	Guar gum	Onion	Rapeseed	Walnut
Carrot	Cow' milk	Gum arabic	Paprika	Rice	Wheat
Cashew	Cumin	Hazelnut	Pea	Rye	

For the following commodities of the table above the results were between 0.5*LOQ and LOQ of the kit. So it cannot be completely excluded that these matrices may provide values above the LOQ in specific cases:

Fish gelatin	Split pea	Turkey	Gum arabic
--------------	-----------	--------	------------

The following cross-reactions were determined:

Conalbumin	< 0.2%
Whole Egg Powder	15%
Lysozyme	< 0.02%
Ovomucoid	< 0.02%

Precision

Intra-assay Precision	3 - 4%
Inter-assay Precision	2 - 3%

Linearity

The serial dilution of spiked samples (white wine, red wine) resulted in a dilution linearity of 94% - 96%.

Recovery

By spiking wine samples (white wine, rosé wine, red wine) with different amounts of ovalbumin, recoveries resulted in 95-115%.

12. REFERENCES

1. Rolland JM, et al. (2008) – Specific and sensitive enzyme-linked immunosorbent assays for analysis of residual allergenic food proteins in commercial bottled wine. J Agric Food Chem, 56(2):349-54
2. Kirschner S, et al. (2009) – Allergenicity of wine containing processing aids: A double-blind, placebo-controlled food challenge. J Inv Clin Imm, 19(3):210-17
3. Weber P, et al. (2007) – Investigation of the allergenic potential of wines fined with various proteinogenic fining agents by ELISA- J Agric Food Chem, 55(8):3127-33

Empfindlichkeit (Ovalbumin)	4 ppb
Wiederfindung	95 - 115%
Inkubationszeit	60 min

1. ALLGEMEINES

Das Ei des Haushuhns (*Gallus gallus*) ist sehr proteinreich und stellt eine bedeutende Nahrungsquelle des Menschen dar. Während die Proteine des Eidotters nur begrenzte Allergenität aufweisen, sind viele Proteine des Eiklars als allergieauslösend bekannt. Neben Lysozym, Ovotransferrin, Ovomucoid und Livetin stellt Ovalbumin die Haupt-fraktion der Eiklarallergene dar. Neben dem klassischen Einsatz in der Nahrungsmittelindustrie, wie zum Beispiel in Backwaren, wird Eipulver oder auch reines Ovalbumin häufig zur Schönung in der Weinindustrie eingesetzt. Der Konsum von Ovalbumin stellt für Allergiker ein kritisches Problem dar. Schon sehr geringe Mengen des Allergens lösen allergische Reaktionen bis hin zum anaphylaktischen Schock aus. Aus diesem Grund müssen Ovalbumin-Allergiker auf den Konsum von Ovalbumin-haltigen Nahrungsmitteln strikt verzichten. Nicht deklarierter Zusatz von Ovalbumin in Lebensmitteln stellt für Allergiker eine Gefahr dar. Kreuzkontaminationen, meist bedingt durch den Produktionsprozess von Nahrungsmitteln, sind ebenfalls problematisch. Seit 1. Juli 2012 verlangt die Europäische Union eine Kennzeichnung für Wein, wenn Milch- oder Ei-Protein im Produktionsprozess verwendet wurde und im Produkt ein Restgehalt von 0,25 mg/L oder höher nachweisbar ist. Um Spuren von Ovalbumin detektieren zu können, bedarf es somit sensitiver Nachweissysteme.

Der **Demeditec Ovalbumin ELISA** stellt ein hoch sensibles Nachweissystem für Ovalbumin dar und ist insbesondere zur Quantifizierung von Ovalbumin-Rückständen in Wein geeignet.

2. TESTPRINZIP

Der **Demeditec Ovalbumin ELISA** basiert auf dem Prinzip des Enzymimmunoassay (EIA). Ein gegen Ovalbumin gerichteter Antikörper ist auf der Oberfläche der Mikrotiterplatte gebunden. Darauf wird die Ovalbumin enthaltende Probe bzw. Standards in die Vertiefungen der Mikrotiterplatte gegeben. Es findet eine Bindung zwischen Antikörper und Ovalbumin statt. Nach 20-minütiger Inkubation bei Raumtemperatur wird die Platte mit verdünntem Waschpuffer gewaschen, um nicht gebundenes Material zu entfernen. Ein Peroxidase-markierter, gegen Ovalbumin gerichteter zweiter Antikörper wird in die Vertiefungen pipettiert. Nach einer weiteren 20-minütigen Inkubation wird erneut gewaschen. Eine Substratlösung wird hinzu pipettiert und 20 Minuten inkubiert, wodurch in den Vertiefungen ein blauer Farbstoff entwickelt wird. Die Farbentwicklung wird durch Zugabe einer Stopp-Lösung beendet, wobei ein Farbumschlag von blau nach gelb stattfindet. Die resultierende Farbe wird spektrophotometrisch bei 450 nm gemessen. Die Ovalbumin-Konzentration ist der Intensität der Färbung direkt proportional.

3. VORSICHTSMAßNAHMEN

Die volle Übereinstimmung mit den folgenden Regeln für eine gute Laborpraxis (GLP) wird zu vertrauenswürdigen Ergebnissen führen:

1. Vor der Testdurchführung müssen alle Reagenzien auf Raumtemperatur gebracht werden (20°C-25°C).
2. Alle Reagenzien sollten vor der Verwendung durch leichtes Kippen oder Schwenken gemischt werden. Die Erzeugung von Schaum sollte dabei vermieden werden.
3. Wenn mit der Testdurchführung einmal begonnen ist, sollten alle nachfolgenden Schritte ohne Unterbrechung und innerhalb der empfohlenen Zeitgrenzen vollzogen werden.
4. Unmittelbar nach Entnahme der Reagenzien sollten die Flaschen wieder mit ihren Stopfen verschlossen werden. Die Verschlüsse dürfen nicht verwechselt werden!
5. Für jede Probe muss eine separate Einmal-Pipettenspitze verwendet werden, um eine Verschleppung bzw. Kreuzkontamination zu vermeiden.
6. Alle Proben und Standards sollten gleichzeitig getestet werden, so dass die Bedingungen für alle identisch sind.
7. Komponenten von verschiedenen Chargen dürfen nicht verwendet oder gemischt werden.
8. Reagenzien dürfen nach der Verfallszeit nicht benutzt werden.
9. Es sollten ständig die Präzision und die Richtigkeit für die Laborgeräte kontrolliert werden, die man für das Testverfahren benutzt (Mikropipetten, Washer, ELISA-Reader, etc.).

4. GESUNDHEITS- UND SICHERHEITSVORSCHRIFTEN

1. Im Laboratorium darf nicht geraucht, gegessen und getrunken werden. Das Pipettieren mit dem Mund ist nicht zulässig.

2. Beim Einsatz von Proben menschlichen Ursprungs müssen Einmalhandschuhe getragen werden.
3. Der Kontakt der Stopp-Lösung mit Haut und Schleimhäuten sollte vermieden werden, da mögliche Reizungen, Verbrennungen oder Vergiftungsgefahr auftreten können. Sollte ein Kontakt entstehen, muss der betroffene Bereich mit ausreichend Wasser abgespült werden.
4. Die Handhabung und die Beseitigung von chemischen Produkten muss nach den Richtlinien für eine gute Laborpraxis (GLP) erfolgen.

5. REAGENZIEN

Der Kit enthält Reagenzien für 96 Bestimmungen. Sie müssen bei 2-8°C gelagert werden. Das Verfallsdatum ist auf jeder Komponente sowie auf der Verpackung angegeben.

1. **SORB MT** Mikrotiterplatte mit 96 Kavitäten (12 Streifen mit je 8 Vertiefungen, einzeln abbrechbar), beschichtet mit Ovalbumin-bindenden Antikörpern.
2. **CAL 1 – 5** Ovalbumin-Standards: 5 Fläschchen mit je 2,0 mL (0, 25, 100, 250, 500 ppb Ovalbumin), rot eingefärbt, gebrauchsfertig.
3. **ENZ CONJ** Konjugat (anti-Ovalbumin-Peroxidase), 15 mL, rot eingefärbt, gebrauchsfertig.
4. **SUB TMB** Substratlösung (TMB), 15 mL, gebrauchsfertig.
5. **STOP SOLN** Stopp-Lösung (0,5 M H₂SO₄), 15 mL, gebrauchsfertig.
6. **SAM DIL 10x** Extraktions- und Probenverdünnungspuffer (TRIS), 2 x 120 mL als 10x-Konzentrat, rot eingefärbt. Gebrauchslösung: 1+9 mit dest. Wasser verdünnen. Der verdünnte Puffer ist bei 4°C mindestens eine Woche haltbar. Falls bei der gekühlten Lagerung Kristalle ausfallen sollten, das Konzentrat 15 Minuten im Wasserbad (37°C) erwärmen.
7. **WASH SOLN 10x** Waschlösung (PBS + Tween 20), 60 mL als 10x-Konzentrat. Gebrauchslösung: 1+9 mit dest. Wasser verdünnen. Der verdünnte Puffer ist bei 4°C mindestens 4 Wochen haltbar. Falls bei der gekühlten Lagerung Kristalle ausfallen sollten, das Konzentrat 15 Minuten im Wasserbad (37°C) erwärmen.
8. Arbeitsanleitung.

6. ZUSÄTZLICH BENÖTIGTE GERÄTE UND REAGENZIEN

Geräte

- 100 - 1000 µL Mikropipetten
- Messkolben
- Analysenwaage
- Mörser, Mixer
- Wasserbad
- Zentrifuge
- Mikrotiterplatten-Photometer (450 nm)
- Ein wiederverschließbarer Plastikbeutel für die Lagerung unbenutzter Streifen

Reagenzien

- bidestilliertes Wasser

7. PROBENVORBEREITUNG

Aufgrund der hohen Gefahr von Kreuzkontaminationen müssen alle verwendeten Arbeitsgeräte wie Spatel, Mörser, Glasgefäße etc. vor und nach jeder Probe **gründlich gereinigt** werden. Es wird empfohlen, die Reihenfolge der Extraktionen festzuhalten, um eventuelle Kreuzkontaminationen durch Vorextrakte besser identifizieren zu können.

Folgende Probenvorbereitung sollte für Wein angewandt werden:

1. 1 mL Weinprobe wird in 19 mL **verdünntem** Extraktionspuffer verdünnt.
2. Für den Test werden pro Kavität 100 µL der Wein-Lösung eingesetzt. Sollte das Ergebnis des Tests oberhalb des Messbereichs liegen, kann die Wein-Lösung gegebenenfalls mit **verdünntem** Extraktionspuffer weiter verdünnt und erneut bestimmt werden.

Folgende Probenvorbereitung sollte für sonstige Proben angewandt werden:

1. Um eine möglichst homogene und repräsentative Probennahme zu gewährleisten, sollten mindestens 5 g Probe in einem Mörser, Schlagmühle etc. fein zermahlen und gut durchmischt werden.
2. Von dieser Mischung wird 1 g entnommen und in 20 mL **verdünntem** Extraktionspuffer suspendiert. Anschließend wird die Suspension für 15 min in einem vorgeheizten Wasserbad bei 60°C inkubiert. Die Proben sollten währenddessen alle zwei Minuten geschüttelt werden, um eine gute Durchmischung zu gewährleisten.
3. Die Proben werden bei mindestens 2500 g für 10 min zentrifugiert. Sollte sich der Überstand anschließend nicht partikelfrei abtrennen lassen, kann gegebenenfalls noch einmal filtriert werden.
4. Für den Test werden pro Kavität 100 µL partikelfreie Lösung eingesetzt. Sollte das Ergebnis des Tests oberhalb des Messbereichs liegen, kann die Lösung gegebenenfalls mit **verdünntem** Extraktionspuffer weiter verdünnt und erneut bestimmt werden.

8. TESTDURCHFÜHRUNG

Der Waschpuffer liegt als 10faches Konzentrat vor und muss vor dem Gebrauch 1+9 mit bidestilliertem Wasser **verdünnt** werden. Die **gebrauchsfertigen** Standards sollten in jedem Fall im Doppelansatz bestimmt werden. Bei größeren Mengen von Proben sollten die Standards einmal vor und einmal nach den Proben pipettiert und der Mittelwert zur Auswertung herangezogen werden. Unter Berücksichtigung von GLP und Qualitätsmanagement ist eine Bestimmung der Proben im Doppelansatz zu empfehlen. Die Testdurchführung verläuft nach dem folgenden Schema:

1. Proben nach Vorschrift vorbereiten.
2. 100 µL **gebrauchsfertige** Standards bzw. vorbehandelte Proben im Doppelansatz in die Vertiefungen der Mikrotiterplatte geben.
3. Platte 20 Minuten bei Raumtemperatur inkubieren.
4. Platte wie folgt dreimal waschen: Vertiefungen der Platte entleeren (auskippen oder absaugen) und 300 µL endverdünnte Waschlösung dazugeben. Nach der dritten Wiederholung Vertiefungen erneut leeren und Flüssigkeitsreste durch Ausschlagen der Mikrotiterplatte auf einem Papiertuch entfernen. Der Waschvorgang ist ein kritischer Schritt. Ungenügendes Waschen führt zu einer geringen Präzision und fälschlicherweise erhöhten Extinktionen.
5. 100 µL Konjugat (anti-Ovalbumin-Peroxidase) in die Vertiefungen pipettieren.
6. Platte 20 Minuten bei Raumtemperatur inkubieren.
7. Vertiefungen wie in Punkt 4 beschrieben waschen.
8. 100 µL Substratlösung zugeben.
9. Platte abdecken und 20 Minuten bei Raumtemperatur im Dunkeln inkubieren.
10. Substrataktion durch Zugabe von 100 µL Stopp-Lösung (0,5 M H₂SO₄) beenden.
11. Nach sorgfältigem Mischen erfolgt die Messung der Extinktion bei 450 nm (eventuell Referenzwellenlänge 620 nm). Die Farbe ist maximal 30 Minuten stabil.

9. BERECHNUNG DER ERGEBNISSE

1. Für jede Probe bzw. jeden Standard wird die gemittelte Extinktion berechnet.
2. Aus den gemittelten Werten der Standardreihe wird graphisch auf halblogarithmischem Millimeterpapier oder über ein EDV-Programm (4-Parameter-Auswertung) eine Eichkurve erstellt. Die Extinktion jedes Standards (y-Achse) wird gegen die Konzentration in ppb (x-Achse) aufgetragen.
3. Mit Hilfe dieser Kurve wird für die gemittelten Extinktionen der Proben der Wert in ppb für Ovalbumin abgelesen. Die ermittelten Konzentrationen beziehen sich direkt auf den Ovalbumingehalt der eingesetzten Probe. Sollte der Extrakt aufgrund eines zu hohen Proteingehalts-Gehalts zusätzlich verdünnt worden sein, muss dies bei der Auswertung berücksichtigt werden.

10. TYPISCHE STANDARDKURVE

Die folgende Tabelle enthält ein Beispiel für eine Standardreihe. Die Werte dieser Standardkurve sind nur als Beispiel bestimmt und dürfen nicht an Stelle der in jedem Test neu zu erstellenden Kurve verwendet werden.

Ovalbumin (ppb)	OD-% von 500 ppb
500	100
250	55
100	26
25	9
0	1

11. TECHNISCHE DATEN

Empfindlichkeit

Die mittlere untere Nachweisgrenze des **Demeditec Ovalbumin Tests** beträgt 4 ppb.

Die untere Bestimmungsgrenze des **Demeditec Ovalbumin Tests** beträgt 25 ppb.

Da jede Matrix einen unterschiedlichen Einfluss auf die LOD haben kann und die Bandbreite der getesteten Matrices begrenzt ist, sollte im Bedarfsfall für jede zu untersuchende Matrix eine spezifische LOD ermittelt werden.

Alternativ können alle Ergebnisse unterhalb der LOQ als quantitativ „< LOQ“ angegeben werden.

Spezifität

Für die folgenden Nahrungsmittel wurde ein negatives Ergebnis und damit keine Kreuz-Reaktivität festgestellt:

Adzuki Bohne	Fischgelatine	Kardamom	Lupine	Pistazie	Sesam
Aprikose	Garnelen, gekocht	Karotte	Macadamia	Porree	Shrimps
Bockshornklee	Garnelen, roh	Kartoffel	Maismehl	Pute	Soja-Lecithin
Bohne, weiß	Gartenkresse	Kichererbse	Mandel (Edeka)	Raps	Sojamehl
Buchweizen	Gerste	Kidneybohne	Marone	Reis	Sonnenblumenkern
Cashew	Gliadin	Kirsche	Meerrettich	Rettich	Thymian
Cayennepfeffer	Guarkernmehl	Kiwi	Milchpulver	Rindergelatine	Tofu
Chia	Gummi arabicum	Knoblauch, frisch	Mohn	Rindfleisch, gekocht	Tomate
Chili	Hafer	Knoblauch, granuliert	Muskatnuss	Rindfleisch, roh	Vollmilch
Cumin	Haselnuss	Kokosnuss	Nelke	Roggen	Walnuss
Dill	Hausenblase	Kümmel	Paprika	Saccharose	Weißkohl
Ente	Huhn	Kürbiskern	Paranuss	Schafsmilch	Weizenkörner
Erbse	Ingwer, frisch	Kurkuma	Pecannuss	Schälererbse	Ziegenmilch
Erdnuss	Ingwer, gemahlen	Lamm	Pfeffer, schwarz	Schweinefleisch	Zimt
Fenchel	Johannisbrotkernmehl	Leinsamen	Pfirsich	Sellerie	Zwiebel
Fisch / Dorsch	Kakao	Linse	Pinienkern	Senf	

Für die folgenden Nahrungsmittel der obenstehenden Tabelle waren die Ergebnisse zwischen 0,5xLOQ und LOQ des Test Kits. Es kann nicht komplett ausgeschlossen werden, dass diese Matrices in Einzelfällen zu Ergebnissen oberhalb der LOQ führen:

Fischgelatine	Pute	Schälererbse	Gummi arabicum
---------------	------	--------------	----------------

Folgende Kreuzreaktionen wurden festgestellt:

Conalbumin		< 0.2%
Volleipulver		15%
Lysozym		< 0.02%
Ovomucoid		< 0.02%

Präzision

Intra-Assay Präzision	3 - 4%
Inter-Assay Präzision	2 - 3%

Linearität

Die schrittweise Verdünnung dotierter Proben über vier Stufen (Weißwein, Rotwein) ergab mittlere Verdünnungslinearitäten von 94 - 96%.

Wiederfindung

Anhand aufgestockter Weinproben (Rotwein, Roséwein, Weißwein) ergaben sich Wiederfindungsra-ten von 95-115%.

12. LITERATUR

1. Rolland JM, et al. (2008) – Specific and sensitive enzyme-linked immunosorbent assays for analysis of residual allergenic food proteins in commercial bottled wine. J Agric Food Chem, 56(2):349-54
2. Kirschner S, et al. (2009) – Allergenicity of wine containing processing aids: A double-blind, placebo-controlled food challenge. J Inv Clin Imm, 19(3):210-17
3. Weber P, et al. (2007) – Investigation of the allergenic potential of wines fined with various proteinogenic fining agents by ELISA- J Agric Food Chem, 55(8):3127-33
4. Yeung JM, et al. (2000) – Determination of egg proteins in food products by enzyme immuno-assay. J AOAC Int, 83(1):139-43
5. Hefle SL, et al. (2001) – Development of a sandwich enzyme-linked immunosorbent assay for the detection of egg residues in processed foodstuff. J Food Prot, 64(11):1812-6

SYMBOLS USED WITH DEMEDITEC ASSAYS

Symbol	English	Deutsch	Française	Espanol	Italiano
	European Conformity	CE-Konformitäts-kennzeichnung	Conforme aux normes européennes	Conformidad europea	Conformità europea
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las Instrucciones	Consultare le istruzioni per l'uso
	In vitro diagnostic device	In-vitro-Diagnostikum	utilisation Diagnostic in vitro	Diagnóstico in vitro	Per uso Diagnostica in vitro
	For research use only	Nur für Forschungszwecke	Seulement dans le cadre de recherches	Sólo para uso en investigación	Solo a scopo di ricerca
	Catalogue number	Katalog-Nr.	Référence	Número de catálogo	No. di catalogo
	Lot. No. / Batch code	Chargen-Nr.	No. de lot	Número de lote	Lotto no
	Contains sufficient for <n> tests/	Ausreichend für "n" Ansätze	Contenu suffisant pour "n" tests	Contenido suficiente para <n> ensayos	Contenuto sufficiente per "n" saggi
	Note warnings and precautions	Warnhinweise und Vorsichtsmaßnahmen beachten	Avertissements et mesures de précaution font attention	Tiene en cuenta advertencias y precauciones	Annoti avvisi e le precauzioni
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeits-datum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Legal Manufacturer	Hersteller	Fabricant	Fabricante	Fabbricante
	Distributed by	Vertrieb durch	Distribution par	Distribución por	Distribuzione da parte di
	Version	Version	Version	Versión	Versione
	Single-use	Einmalverwendung	À usage unique	Uso único	Uso una volta